



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Florida Highway Patrol

Instrument Serial Number: 80-000227

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument <input checked="" type="checkbox"/> Instrument Set Up Verified	<input checked="" type="checkbox"/> Agency Inspection Notification <input type="checkbox"/> Agency Contact <input type="checkbox"/> Other: _____ Date of Notification/Contact: 07/30/2018 Agency Inspector: Trooper Joseph Farley	The following records were audited: <input checked="" type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____

Comments:

Trooper Joseph Farley was contacted on July 30th, 2018 to notify him that the Alcohol Reference Solutions used during agency inspections from November 2017 through July 2018 are not approved by the FDLE for use. Trooper Farley was informed that this is a Rules violation.

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces

Comments:

Trooper Farley has another instrument in his possession which he performs inspections on with the same solutions and was informed to use approved solutions with orange labels per the current format. A department inspection was conducted on this instrument upon return from FHP.

Action:	
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC	Comments: The instrument appears to have been working accurately and reliably regardless of the unapproved nature of the solutions used.
<input type="checkbox"/> Remain in Evidentiary Use <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use	
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	

*Joseph Farley*

Signature of Alcohol Testing Program Staff Member

7/31/18

Date

Original - FDLE

Copy - Agency