



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Okaloosa County Sheriff's Office Instrument Serial Number: 80-000739

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Records Audit items.

Comments:

A.I. reported a failed inspection. The 0.200g/210L measurement was low. As all the other measurements were in tolerance, a simulator leak was the most probable cause. He was authorized to repeat the inspection using a different simulator.

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, and various supplies.

Comments:

Action, Comments. Includes checkboxes for Instrument Complies with Chapter 11D-8, FAC and various evidentiary use options.

Signature of Alcohol Testing Program Staff Member (Handwritten signature)

Date (Handwritten: 2/27/2018)

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency