



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: New Smyrna Beach Police Department Instrument Serial Number: 80-001152

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Records Audit items.

Comments: .A.I. got an "interferent detect" during the alcohol free test. He cleaned the simulator. He then re-ran the inspection, which turned out problem free.

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, and various supplies.

Comments:

Action, Comments. Includes checkboxes for Instrument Complies with Chapter 11D-8, FAC and other actions.

Signature of Alcohol Testing Program Staff Member (Handwritten: Patrick Murphy)

Date (Handwritten: 9/4/2018)

Original - FDLE Copy - Agency



Florida Department of  
Law Enforcement

**DEPARTMENT INSPECTOR  
FIELD NOTES**

Agency: New Smyrna Beach P.D.

Instrument Serial Number: 80-001152

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____  <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument  <input type="checkbox"/> Instrument Set Up Verified	<input checked="" type="checkbox"/> Agency Inspection Notification <input type="checkbox"/> Agency Contact <input type="checkbox"/> Other _____  Date of Notification/Contact: <u>4/29/2018</u>  Agency Inspector: <u>Sgt. Jason Reve</u>	The following records were audited: <input type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis  <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____

Comments:

A.I. reported failed inspection due to operator error. Instrument passed during second attempt.

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces

Comments:

Action:	Comments:
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC	
<input type="checkbox"/> Remain in Evidentiary Use <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use	
<input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	

*Patrick Murphy*

Signature of Alcohol Testing Program Staff Member

5/5/2018

Date

Original - FDLE

Copy - Agency