80-006633 KMB

FDLE
Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: FL Highway Patrol Instrument Serial Number: 80-006633

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☑ Agency Inspection Notification ☑ Agency Contact ☐ Other _____ Date of Notification/Contact: 06/25/2018 18:21 Agency Inspector: Alain Hernandez-Ginarte	The following records were audited: ☐ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Instrument Registration ☐ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:
Al Hernandez called to notify me that an inspection on 80-006633 cannot be conducted because the instrument will not come into ready mode due to EEPROM diagnostic fails at the start-up. I instructed him to box up the instrument and ship it to me at ATP Tallahassee.

Instrument/Area:	Equipment:	Supplies:
☐ Clean/Dry ☐ Secure ☐ Limited Access ☐ Other: _____	☐ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☐ Distilled/Deionized Water ☐ Mouth Alcohol Solution ☐ Acetone Stock Solution ☐ Alcohol Reference Solution ☐ Dry Gas Standard ☐ Mouth Pieces

Comments:

Action:	
☐ Instrument Complies with Chapter 11D-8, FAC ☑ Instrument Does Not Comply with Chapter 11D-8, FAC	Comments:
☐ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☑ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

Signature of Alcohol Testing Program Staff Member Date: 6/26/18

Original – FDLE Copy – Agency