



Florida Department of Law Enforcement

**DEPARTMENT INSPECTOR
FIELD NOTES**

Agency: Florida Highway Patrol Instrument Serial Number: 80-006764

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
<input type="checkbox"/> Registration <input type="checkbox"/> Request for Registration Attached <input type="checkbox"/> Annual <input type="checkbox"/> Inspection After Repair <input type="checkbox"/> Other: _____ <input type="checkbox"/> Barometric Pressure: _____ Gauge _____ Instrument <input type="checkbox"/> Instrument Set Up Verified	<input checked="" type="checkbox"/> Agency Inspection Notification <input type="checkbox"/> Agency Contact <input type="checkbox"/> Other _____ Date of Notification/Contact: <u>07/30/2018</u> Agency Inspector: <u>Trooper Joseph Farley</u>	The following records were audited: <input checked="" type="checkbox"/> Agency Inspection Reports <input type="checkbox"/> Breath Test Instrument Repair/Maintenance Records <input type="checkbox"/> Instrument Registration <input type="checkbox"/> Dry Gas Standard Certificate(s) of Analysis <input type="checkbox"/> Other: _____ <input type="checkbox"/> Other: _____

Comments:
Trooper Joseph Farley was contacted on July 30th, 2018 to notify him that the Alcohol Reference Solutions used during agency inspections in 2017 and 2018 are not approved by the FDLE for use. Trooper Farley was informed that this is a Rules violation and corrective measures are

Instrument/Area:	Equipment:	Supplies:
<input type="checkbox"/> Clean/Dry <input type="checkbox"/> Secure <input type="checkbox"/> Limited Access <input type="checkbox"/> Other: _____	<input type="checkbox"/> Proper Number of Simulators <input type="checkbox"/> Checked Simulators for Air Leak Resistant Seal and Proper Temperature <input type="checkbox"/> Class A Glassware	<input type="checkbox"/> Distilled/Deionized Water <input type="checkbox"/> Mouth Alcohol Solution <input type="checkbox"/> Acetone Stock Solution <input checked="" type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard <input type="checkbox"/> Mouth Pieces

Comments:
required immediately. Trooper Farley and his co-workers perform inspections on the three (3) instruments assigned together. The same solutions are used on all of their instruments. Farley was instructed to notify his co-workers and do an inspection with approved solutions immediately.

Action:	
<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC	Comments: The instrument appears to have been working accurately and reliably regardless of the unapproved nature of the solutions used.
<input type="checkbox"/> Remain in Evidentiary Use <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remove from Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use	
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	

[Handwritten Signature]

Signature of Alcohol Testing Program Staff Member

7/31/18

Date

Original - FDLE

Copy - Agency