FDLE
Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Tallahassee PD Instrument Serial Number: 80-006926

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☑ Agency Inspection Notification ☐ Agency Contact ☐ Other _____ Date of Notification/Contact: 2/28/2018 Agency Inspector: Terry Mahan	The following records were audited: ☐ Agency Inspection Reports ☐ Breath Test Instrument Repair/Maintenance Records ☐ Instrument Registration ☐ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:
A.I. reports a failed inspection. He hadn't connected the printer and received a printer error when he tried to print. He then backed into the program set. He was authorized to repeat the inspection.

Instrument/Area:	Equipment:	Supplies:
☐ Clean/Dry ☐ Secure ☐ Limited Access ☐ Other: _____	☐ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☐ Class A Glassware	☐ Distilled/Deionized Water ☐ Mouth Alcohol Solution ☐ Acetone Stock Solution ☐ Alcohol Reference Solution ☐ Dry Gas Standard ☐ Mouth Pieces

Comments:

Action:	
☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	Comments:
☐ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

Patrick Murphy — Signature of Alcohol Testing Program Staff Member

2/28/18 — Date

Original – FDLE Copy – Agency