

**Bell, Danielle**

---

**From:** John Sommersdorf <John.Sommersdorf@leesburgflorida.gov>  
**Sent:** Friday, March 09, 2018 4:16 AM  
**To:** Bell, Danielle  
**Subject:** RE: Agency inspection

It was a long night. Simulator seal looked fine. When I reran I left jar loose on acetone so it required a second time.

I had replaced all of the orings on the hoses last month so I assumed I had small ones causing leaks. It would have been easier to check jar seal.

**From:** John Sommersdorf  
**Sent:** Friday, March 9, 2018 2:35 AM  
**To:** 'DanielleBell@fdle.state.fl.us' <[DanielleBell@fdle.state.fl.us](mailto:DanielleBell@fdle.state.fl.us)>  
**Subject:** Agency inspection

I had the .080 out of tolerance, I thought it was just orings so I switched hoses with another jar and reran test. Still had a low reading so I tightened jar extra tight and got two good readings. I marked that test not in compliance and I am rerunning test since I believe I fixed the problem. When I finish I will check to see if simulator jar seal is bad, I shouldn't have to crank it down that much.

Sommersdorf  
Leesburg Police Department.

---

Please Note: Florida has a very broad public records law. Most written communications to or from government officials regarding government business are public records available to the public and media upon request. Your e-mail communications may therefore be subject to public disclosure.

Think before you print.