

Platt, Shayla

995
SP

From: James Scheiner <jscheiner@cityofcocoabeach.com>
Sent: Thursday, June 14, 2018 4:56 PM
To: Platt, Shayla
Cc: Diana Custer; James Carroll; Kelly Swartz
Subject: RE: 995
Attachments: SKM_C224e18061416550.pdf

Shayla,

Attached is my amended FDLE form 40. If there is any further changes just advise. The test was started as soon as the Sim was up to temperature. More time was needed with the Sim running for the test results to be within tolerance. Moving forward I will be spending more time between each simulator test to ensure this issue does not arise again. My agency has also purchased more glassware so I can utilize all of our simulators.

James Scheiner ID16
Cocoa Beach Police Department
321-868-3251
20 South Orlando Ave
Cocoa Beach FL 32931

From: Platt, Shayla [<mailto:ShaylaPlatt@fdle.state.fl.us>]
Sent: Thursday, June 14, 2018 10:56 AM
To: James Scheiner
Subject: 995

Good Morning,

Please see the attached information and forward requested information.

Thanks,

Shayla Platt

Department Inspector, Alcohol Testing Program
Florida Department of Law Enforcement
2331 Phillips Rd.
Tallahassee, FL 32308
(850) 617-1282 (office)
(850) 728-4200 (cell)
(850) 921-3787 (fax)

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.