

DANIELLE - HERE IS A COPY OF THE ORIGINAL FORM 40.

\* COULD YOU PLEASE CHECK THIS INSTRUMENT OVER, IT PASSED AFTER REPEATING "AIR" + .050 !!  
FREE

I HAVE NEVER HAD AN AMBIENT FAIL ON OVER 1800 INSPECTIONS I HAVE COMPLETED.

\* THIS INSTRUMENT WAS ONE OF THREE THAT FHP HAD MANY PROBLEMS WITH ... THIS ONE WAS UTILIZED 5 TIMES BEFORE MOVING TO ANOTHER INSTRUMENT ...

ALL 7 INSTRUMENTS WERE UPLOADED WITH NO PROBLEMS.

KEEP ME POSTED ON WHAT YOU FIND WRONG, IF ANYTHING.

THANKS - DINO ☺

80-001003

RECEIVED 

OCT 30 2018

FDLE  
Alcohol Testing Program

SP

11/6/18  


**Bell, Danielle**

---

**From:** Deane II, Malcolm <mdeane@pcsonet.com>  
**Sent:** Thursday, October 25, 2018 7:52 PM  
**To:** Bell, Danielle  
**Subject:** # 1003 Instrument

Hello D. B. – I am shipping up to you # 1003 for you to perform a general inspection. This is the instrument that had an operator error happen to it by an outside agency. I included a copy of the Form 40 with a note on the back Thank you in advance for your efforts – Dino :) P.S. – could you send back some Acetone Stock ... I have one left with expiration of Feb. 2019 - thanks