

## Coxetter, Quin

---

**From:** Graham, Thomas  
**Sent:** Friday, December 28, 2018 9:57 AM  
**To:** thammonds@polksheriff.org  
**Cc:** Coxetter, Quin; kwiegert@polksheriff.org  
**Subject:** December AI for 80-001055  
**Attachments:** Polk County December AI.pdf

Deputy Hammonds,

Attached is a copy of your December 2018 Agency Inspection. As we discussed on the telephone, your printer port is loose and you are unable to connect your external printer to the Intoxilyzer (80-001055). You said you will ship me the instrument and I will determine if it needs to be sent to a repair facility. I will follow up with you should your instrument need to be sent to repair.

Thank you for contacting us. Have a happy New Year!

TJ Graham, D-ABFT-FA  
Quality Assurance Manager  
Alcohol Testing Program  
Florida Department of Law Enforcement  
Office:850-617-1285  
Cell: 850-445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.