

From: [Bell, Danielle](#)
To: [Graham, Thomas](#)
Subject: FW: Intox Issue
Date: Thursday, January 03, 2019 9:03:02 AM
Attachments: [image001.png](#)
[image003.png](#)

Documentation of the conversations regarding the failed inspection of 80-001061 in December 2018.

Thanks,

Danielle

Danielle M. Bell

Sr. Mgmt. Analyst Supervisor
Office of Field Services
Systems Administration Unit
Florida Department of Law Enforcement
(850) 410-8516 (office)
(850) 688-7553 (cell)

From: Bell, Danielle
Sent: Tuesday, January 01, 2019 6:20 PM
To: Marotta, Mike
Subject: Re: Intox Issue

Nope. You did everything right! Thanks! TJ Graham is going to do the inspection and get it back to you asap. He's great and you'll like him. He's technically the Quality Manager, but he's performing DI duties while they search for my replacement. He is super knowledgeable and responsive.

TJ Graham
ThomasGraham@fdle.state.fl.us
(850) 617-1285

P.S. I'm always around. But more for social calls now.

Thanks

Danielle

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Marotta, Mike" <mymarotta@mygulfport.us>
Date: 1/1/19 6:09 PM (GMT-05:00)
To: "Bell, Danielle" <DanielleBell@fdle.state.fl.us>
Subject: RE: Intox Issue

Hello,

I was able to get the intox to pass today's inspection. I packed it up to ship up for its annual inspection. It should be mailed out tomorrow. Is there anything else I need to do?

Sgt. Michael Marotta
Gulfport Police Dept.
2401 53rd St S.
Gulfport, Fl. 33707
Dispatch: (727) 582-6177
Office: (727) 893-1052

Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.



From: Marotta, Mike
Sent: Monday, December 31, 2018 1:04 PM
To: Bell, Danielle <daniellebell@fdle.state.fl.us>
Subject: Re: Intox Issue

That sounds like a good idea.

Sgt. Michael Marotta
Gulfport Police Dept.
2401 53rd St S.
Gulfport, Fl 33707
Dispatch: (727) 582-6177
Office: (727) 893-1052

Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

From: Bell, Danielle <DanielleBell@fdle.state.fl.us>
Sent: Monday, December 31, 2018 11:00:49 AM
To: Marotta, Mike
Subject: Re: Intox Issue

Fine by me. If you so desired, you could get the good inspection then send it up to ATP for its annual and be done for the year.

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Marotta, Mike" <mymarotta@mygulfport.us>
Date: 12/31/18 10:27 AM (GMT-05:00)
To: "Bell, Danielle" <DanielleBell@fdle.state.fl.us>
Subject: Re: Intox Issue

I am off until tomorrow. I'll be in at 1600. If you've good with me trying again then I'll just keep it out of service until that time.

Sgt. Michael Marotta
Gulfport Police Dept.
2401 53rd St S.
Gulfport, FL 33707
Dispatch: (727) 582-6177
Office: (727) 893-1052

Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

From: Bell, Danielle <DanielleBell@fdle.state.fl.us>
Sent: Monday, December 31, 2018 2:26:51 AM
To: Marotta, Mike
Subject: Re: Intox Issue

You up for trying again tomorrow? Let's heat up just a .20 solution and do some stability tests to confirm whether or not it actually was the sim. I'm assuming both times were too low. If we can get a sim up to the correct temp and without leaks to come into range, we'll heat the rest of the sims and do an inspection. If you come on shift by 8 and get the 0.20 heating asap, you'll have enough

time to get it online before midnight. If you come on shift earlier than that, you'll have more cushion for testing. You have 4 sims, correct? If you're game, call me when you come on shift tomorrow and we can discuss. :)

Thanks,
Danielle
New cell: 850 688 7553

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Marotta, Mike" <mymarotta@mygulport.us>
Date: 12/30/18 10:15 PM (GMT-05:00)
To: "Bell, Danielle" <DanielleBell@fdle.state.fl.us>
Subject: Intox Issue

Hello,

Hope you're doing well. I know you're in a new position, however, I wasn't sure who I should direct my issue to. I could not get my .20 solution within range either time I attempted it, so it did not pass the inspection. It may just be a situation where the solution was not warm enough, however, I waited over 30 minutes in the first sim and another 15 minutes when I switched sims. Please let me know how I need to proceed. I took the instrument out of service for now.

Thanks,

Sgt. Michael Marotta
Gulfport Police Dept.
2401 53rd St S.
Gulfport, Fl. 33707
Dispatch: (727) 582-6177
Office: (727) 893-1052

Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

