

## Coxetter, Quin

---

**From:** Coxetter, Quin  
**Sent:** Friday, December 28, 2018 12:59 PM  
**To:** 'jstone@vcso.us'  
**Subject:** December 2018 Agency Inspection 80-001130  
**Attachments:** 80-001130 AI Dec 2018.pdf

Good afternoon Mr. Stone,

Per our phone conversation I understand that you had a printer error when trying to print your Agency Inspection for December 2018 on instrument

80-001130. Per your request I have attached the electronic copy of your Agency Inspection for your records. If you need anything further please do not hesitate to contact me directly.

Thank you,

Quin Coxetter  
Alcohol Testing Program  
Florida Department of Law Enforcement  
(850) 617-1278 (office)

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.

Serial Number: 80-001130

Time of Inspection: 12:26

Date of Inspection: 12/28/2018

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 644476 Exp: 12/01/2019
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.201	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JIMMIE N STONE

Signature and Printed Name

12/28/2018

Date