

Bell, Danielle

From: Bell, Danielle
Sent: Monday, April 16, 2018 8:40 AM
To: 'Andy Crogan'
Subject: RE: Agency inspections
Attachments: Form 40-Agency Inspection Report - Intox 8000.pdf

Start over. Document on a blank form 40 (attached) what happened with anything you can remember from the first test and restart. Oh, send me a copy of any handwritten or typed Form 40s.

Thanks,

Danielle

Danielle M. Bell

Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

From: Andy Crogan [<mailto:acrogan@ladylake.org>]
Sent: Monday, April 16, 2018 7:24 AM
To: Bell, Danielle
Subject: Agency inspections

Danielle,

I hade an operator error during my monthly. When I unplugged the .200 sim I also unplugged the instrument. Any way to save that, or just start over?

Thanks,
Andrew Crogan
Lady Lake Police Department
(352)636-5689

Sent from my Verizon, Samsung Galaxy smartphone

Fla. Stat. 668.6076 "Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing."