

Bell, Danielle

From: Tom Lokietek <TLokietek@auburndalefl.com>
Sent: Tuesday, February 20, 2018 5:10 PM
To: Bell, Danielle
Subject: RE: Intox shipped

No, there were none completed.

From: Bell, Danielle [<mailto:DanielleBell@fdle.state.fl.us>]
Sent: Tuesday, February 20, 2018 12:43 PM
To: Tom Lokietek
Subject: RE: Intox shipped

Good morning,

It sounds like the battery that controls the clock is down. Do you know if there were any subject tests done during January after the monthly inspection?

Thanks,

Danielle

Danielle M. Bell
Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

From: Tom Lokietek [<mailto:TLokietek@auburndalefl.com>]
Sent: Monday, February 19, 2018 11:34 PM
To: Bell, Danielle
Subject: Intox shipped

Good morning,

We had an issue with our Intoxilyzer. Detective Oros completed the monthly inspection for January. When Corporal Baker attempted to use it in February he received a "Disabled mode monthly AI expired" reading. When I turned the instrument on the date displayed as 01/01/2020. It has been mailed and you should be receiving it shortly. Thanks.

Sergeant Thomas J. Lokietek
Law Enforcement Division
Auburndale Police Department
2 Bobby Green Plaza
Auburndale, FL 33823
(Office)863-965-6363
(Cell)863-528-7939