

80-001732 RMB



PINELLAS PARK POLICE DEPARTMENT
SPECIAL OPERATIONS SECTION
MEMORANDUM



TO: Department Inspector Danielle Bell
FROM: Ofc Lawrence R Kolbicka II
RE: Intoxilyzer 8000
DATE: 11/2/2018

RECEIVED
NOV 02 2018
FDLE
Alcohol Testing Program

On 11/2/2018 I conducted an agency inspection for Intoxilyzer 1732. I had to abort the sequence after I connected what I thought was the incorrect simulator. After aborting the sequence, the instrument went back to the home screen and did not ask if the instrument was in compliance or if I wanted to print. I attempted to locate the record in the recall and it was not there.

I conducted an inspection immediately after and there were no issues.

Respectfully Submitted,

Ofc Lawrence Kolbicka II