



NEW SMYRNA BEACH POLICE DEPARTMENT

246 INDUSTRIAL PARK AVE | NEW SMYRNA BEACH, FL 32168

May 1, 2018

Sergeant Jason R. Reve
246 Industrial Park Avenue
New Smyrna Beach, FL 32168

Patrick Murphy, Ph.D.
2331 Phillips Road
Tallahassee, FL 32308
To Whom It May Concern:

RECEIVED
MAY 03 2018
FDLE
Alcohol Testing Program

Mr. Murphy,

On April 29, 2018, while conducting the agency inspection for Intoxilyzer 8000 SN#80-00223, I experienced several "Ambient Fail" indications during the mouth alcohol/alcohol free samples. I attempted to troubleshoot the problem and was unable to alleviate the persistent ambient fail condition.

Per your recommendation I attempted an agency inspection again the following day, April 30, 2018. Once the .05 solution sample was introduced the instrument produced a "fail to purge". Again I attempted to troubleshoot and was meant with an "interferent detect" and "ambient fail" on subsequent attempt.

As discussed, I am sending the instrument back to you at the FDLE Alcohol Testing Program for further troubleshooting and repair.

Regards,

Sergeant Jason R. Reve
jreve@cityofnsb.com
Office: (386) 402-7743
Cell: (386) 576-6830