

Bell, Danielle

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Friday, September 07, 2018 1:19 AM
To: Bell, Danielle
Cc: George Maroney; Jacob Beekman
Subject: Agency Inspection Sept 2018
Attachments: 06 Sept 2018 montly intox inspec.pdf

Danielle,

Before I forward this out, I want you to take a look and see if I need to annotate my corrections any further.

Had a bad bottle of 0.05 solution.

Ref to instrument 80-004587, the 0.05g/210L Test – I received a 05: Control outside tolerance. Solution read 0.000. Solution bottle 0819.

Solution was changed (bottle 0820) but for instrument 80-004587, the first reading was outside tolerance. I only allowed to sit for about 10 mins and believed it was not up to temp for that reading.

I went to print and hit wrong key and restarted. I aborted the sequence to re-start the test. After allowing additional wait time for solution to come up to temp. re-test was complete and within standards.

Same simulator with same solutions, bottle 0819 was used on instrument 80-000968, with the same reading of 0.000. Changed solution (bottle 0820).

Retested instrument 80-000968 with results within standards

Thanks,

Dan

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Bell, Danielle

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Wednesday, August 08, 2018 11:26 PM
To: Bell, Danielle; 'McKnight, Jamie (JMcknight@sao9.org)'; 'SherryForcier@flhsmv.gov'; 'Atkinson, Karsona'
Cc: George Maroney; Jacob Beekman
Subject: Monthly Intox Inspections for Aug 2018
Attachments: 08 Aug 2018 montly intox inspec.pdf

ALCON,

The Agency Inspection Report was conducted on 08/08/2018 on the following seven (7) instruments; 80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-000968, the 0.08g/210L Test. I received a Control outside tolerance. I reset the lid on the simulator and re-tested. The re-test results are within standards. The inspection was re-tested, not re-started.

Ref to instrument 80-003937, the 0.05g/210L Test – I received a 05: Control outside tolerance. I failed to connect the simulator. I connected the simulator and re-tested. The inspection was re-tested, not restarted. The re-test results are within standards.

Ref to instrument 80-005935, A F / M A Test – I received Ambient fail. I removed the mouth piece from the hose and re-tested. The inspection was re-tested, not restarted. The re-test results are within standards.

Instrument 80-003935 was returned by CMI upon completion of the Annual inspection, inspected and returned to service.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message

Bell, Danielle

80-804587 ~~1018~~

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Wednesday, June 06, 2018 1:19 AM
To: Bell, Danielle; 'McKnight, Jamie (JMcKnight@sao9.org)'; 'SherryForcier@flhsmv.gov'; 'Atkinson, Karsona'
Cc: George Maroney; Jacob Beekman
Subject: RE: Monthly Intox Inspections for May 2018
Attachments: 05 June 2018 montly intox inspec.pdf

Follow Up Flag: Follow up
Flag Status: Flagged

ALCON,

Agency Inspection Report was conducted on 06/05/2018 on the following six (6) instruments; 80-000966, 80-000968, 80-001715, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-003937, Ref to the 0.08 g/210L Dry Gas Std Test – I received a 08: Control Outside Tolerance. I failed to connect the Dry Gas Line Prior to start of test. I connected the line and re-tested. The re-test results are within standards. The inspection was re-tested, not re-started.

Ref to instrument 80-005935, Ref to the Int Det Test – I failed to connect the sim. When asked to retest, I selected N in error, resulting in a restart. Inspection was restarted. Test results are within standards.

Ref to instrument 80-000968, Ref to the A F / M A Test – I failed to provide the requested sample within the allotted time. When asked to retest, I selected N in error, resulting in a restart. The inspection was restarted. In ref to the 0.08g/210L test, I attached the wrong (0.20g/210L solution) simulator to the instrument. I replaced it with the proper simulator and re-tested. The inspection was re-tested, not restarted. The re-test results are within standards.

Instrument 80-003935 is still out at CMI for Annual inspection.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

Bell, Danielle

From: Dan Lyons <dlyo@OSCEOLA.ORG>
Sent: Tuesday, March 27, 2018 12:16 AM
To: Bell, Danielle; 'McKnight, Jamie (JMCKnight@sao9.org)'; 'SherryForcier@flhsmv.gov'; 'Atkinson, Karsona'
Cc: George Maroney
Subject: Monthly Intox Inspections for March 2018
Attachments: 26 March 2018 monthly intox inspec.pdf

ALCON,

Agency Inspection Report conducted 03/26/2018 on the following seven (7) instruments; 80-000966, 80-000968, 80-001715, 80-003935, 80-003937, 80-004587 and 80-005935.

Ref to instrument 80-005935, Ref to the 0.08/210L Test – I received a 08: Control Outside Tolerance. I reset the top of the simulator and re-tested. The re-test results are with in standards. The inspection was re-tested, not re-started.

Instruments 80-000968 and 80-003937 have been returned from FDLE for Annual inspections, and upon completion of the March 2018 monthly inspections, were placed back into service.

Instruments 80-00393 has been sent off to FDLE for Annual inspection.

All intox inspections are complete and in compliance with 11D-8.

V/R

Dan Lyons
Osceola County Sheriff's Office
Spec Ops/Traffic/DUI Enforcement Unit
Agency 407-348-1100
Agency Cell 407-269-1980
Cell 321-960-2066

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.