

Bell, Danielle

From: Hernandez-Ginarte, Alain <AlainHernandez-Ginarte@flhsmv.gov>
Sent: Friday, September 28, 2018 3:58 AM
To: Bell, Danielle
Cc: Campbell, Joshua
Subject: 80-006638 AND 80-006046

Good morning Danielle,

I did the inspections last night and like every month, at least 1 of the instruments gave me a problem.

80-006638: Continued to give me ambient failure and I was unable to proceed with the monthly inspection on it. I tried 3 times on AF/MA, and continued to do Ambient Failure on the air blanks. Good thing is that this instrument hasn't been used for the month of September. I will send it to you ASAP when I get a chance.

80-006046: Prints out gibberish on the printer (same as last month). I already faxed it to ya so if you could email me the inspection report I would appreciate it!

Other than that, all the other intoxes passed the inspection. Thank you.

Trooper Alain Hernandez
Florida Highway Patrol
Troop C- Special Operations District / DUI Enforcement
Drug Recognition Expert – IACP#26438
DUI Instructor
11305 North Mckinley Drive
Tampa, FL. 33612
Office: 813.558.1800
AlainHernandez-Ginarte@flhsmv.gov
<http://www.flhsmv.gov/fhp>