

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO
Time of Inspection: 10:09

Date of Inspection: 06/12/2019

Serial Number: 80-000739
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:754250 Exp: 04/03/2020
0.000	0.046	0.077	0.187 / 0.196	0.080
0.000	0.048	0.080	0.197 / 0.197	0.080
0.000	0.049	0.080	0.198 / 0.197	0.080

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. *Retightened Lid*
Amended copy David L Johnson

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David L Johnson

DAVID L JOHNSON *632/132*

Signature and Printed Name

06/12/2019
Date