

PO-000770
BK
2/20/19

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 17:55

Date of Inspection: 01/31/2019

Serial Number: 80-000770
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2018 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 03417080A1 Exp: 03/05/2019 |
|----------------------------|--|--|--|--|
| 0.000 | 0.050 | 0.079 | 0.193 | 0.076 |
| 0.000 | 0.050 | 0.080 | 0.193 | 0.076 |
| 0.000 | 0.051 | 0.080 | 0.193 | 0.076 |

Number of Simulators Used: 4

Remarks: *Amendment on 0.05 Alcohol Solution Date
Correct Date is 2/06/2020 Done on the 20th Day
of February 2019 W.C. Curry*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

W.C. Curry

Signature and Printed Name WILLIAM C CURRY

01/31/2019
Date