

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 14:36

Date of Inspection: 05/05/2019

Serial Number: 80-000781

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:1076482 Exp: 01/15/2022 <i>See comments</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1076482 Exp: 01/15/2022
0.000	0.049	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079

Number of Simulators Used: 5

Remarks:

LOT#201802G EXP 02-22-20

JMB

*lot# 201802G
EXP 02-22-2022
JMB*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan M Barge
Signature and Printed Name

SUSAN M BARGE

05/05/2019
Date