

80-000797
TBK

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 18:34

Date of Inspection: 07/31/2019

Serial Number: 80-000797
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 00919080A3 Exp: 03/05/2021
0.000	0.043 / 0.047	0.077	0.188 / 0.193	0.082
0.000	0.043 / 0.048	0.079	0.191 / 0.194	0.082
0.000	0.044 / 0.048	0.079	0.191 / 0.194	0.082

Number of Simulators Used: 4

AMENDMENT

Remarks:
 f05 CHANGED SIMULATOR HEATED UP AND RE CONDUCTED TEST
 (0.20 Simulator was still heating upon first test)
 e. CHANGED SIMULATOR

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry

Signature and Printed Name: WILLIAM C CURRY

07/31/2019
Date