80-000812 BK

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 11:42 Date of Inspection: 12/29/2019 Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.197 | 0.082 |
| 0.000 | 0.048 | 0.079 | 0.197 | 0.082 |
| 0.000 | 0.048 | 0.079 | 0.197 | 0.082 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Q 1193 BRAD R COLLITO
Signature and Printed Name

12/29/2019
Date

*FDLE/ATP Form 40 – March 2004*

80-000812
BK

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 17:20 Date of Inspection: 11/26/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.078 | 0.195 | 0.082 |
| 0.000 | 0.049 | 0.080 | 0.197 | 0.082 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.082 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1193 BRAD R COLLITO
Signature and Printed Name

11/26/2019
Date

*FDLE/ATP Form 40 – March 2004*

80-000812
BK

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 09:45 Date of Inspection: 10/28/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.197 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1193 BRAD R COLLITO
Signature and Printed Name

10/28/2019
Date

*FDLE/ATP Form 40 – March 2004*

80-000812
BK

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Serial Number: 80-000812
Time of Inspection: 22:44
Date of Inspection: 09/15/2019
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.197 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.081 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1493 BRAD R COLLITO
Signature and Printed Name

09/15/2019
Date

*FDLE/ATP Form 40 – March 2004*

RECEIVED
AUG 28 2019
FDLE
Alcohol Testing Program

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 11:47 Date of Inspection: 08/25/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.196 | 0.080 |
| 0.000 | 0.048 | 0.080 | 0.197 | 0.080 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Co 193 BRAD R COLLITO
Signature and Printed Name

08/25/2019
Date

*FDLE/ATP Form 40 – March 2004*

# Florida Department of Law Enforcement
# Alcohol Testing Program

7/22/19
JC

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 13:31 Date of Inspection: 07/21/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.197 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |

Number of Simulators Used: 5

Remarks:
A F / M A: Range Exceeded RESIDUAL FROM PREVIOUS TEST. BC1193

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B C 1193 BRAD R COLLITO
Signature and Printed Name

07/21/2019
Date

*FDLE/ATP Form 40 – March 2004*

# Florida Department of Law Enforcement
# Alcohol Testing Program

6/21/19

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 15:17 Date of Inspection: 06/20/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A2 Exp: 03/05/2021 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.048 | 0.080 | 0.199 | 0.081 |
| 0.000 | 0.048 | 0.080 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Q 1193 BRAD R COLLITO
Signature and Printed Name

06/20/2019
Date

*FDLE/ATP Form 40 – March 2004*

# Florida Department of Law Enforcement
# Alcohol Testing Program

5/9/19 JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 13:39 Date of Inspection: 05/09/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34917080A3 Exp: 02/05/2020 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.078 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1193 BRAD R COLLITO
Signature and Printed Name

05/09/2019
Date

# Florida Department of Law Enforcement
# Alcohol Testing Program

4/16/19 JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 19:02 Date of Inspection: 04/14/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34917080A3 Exp: 02/05/2020 |
|---|---|---|---|---|
| 0.000 | 0.048 | 0.078 | 0.197 | 0.080 |
| 0.000 | 0.048 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.048 | 0.079 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1193 BRAD R COLLITO
Signature and Printed Name

04/14/2019
Date

*FDLE/ATP Form 40 – March 2004*

# Florida Department of Law Enforcement
# Alcohol Testing Program

3/12/19 JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 20:01 Date of Inspection: 03/11/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34917080A3 Exp: 02/05/2020 |
|---|---|---|---|---|
| 0.000 | 0.047 | 0.077 | 0.191 | 0.081 |
| 0.000 | 0.047 | 0.077 | 0.191 | 0.081 |
| 0.000 | 0.048 | 0.078 | 0.192 | 0.082 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B C 1193 BRAD R COLLITO
Signature and Printed Name

03/11/2019
Date

# Florida Department of Law Enforcement
# Alcohol Testing Program

2/12/19
JO

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 03:35 Date of Inspection: 02/12/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34917080A3 Exp: 02/05/2020 |
|---|---|---|---|---|
| 0.000 | 0.049 | 0.077 | 0.197 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.197 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B C 1193 BRAD R COLLITO
Signature and Printed Name

02/12/2019
Date

*FDLE/ATP Form 40 – March 2004*

# Florida Department of Law Enforcement
# Alcohol Testing Program

2/7/19

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 20:13 Date of Inspection: 01/09/2019
Serial Number: 80-000812
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34917080A3 Exp: 02/05/2020 |
|---|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B Collito 1183 BRAD R COLLITO
Signature and Printed Name

01/09/2019
Date

*FDLE/ATP Form 40 – March 2004*