

Florida Department of Law Enforcement Alcohol Testing Program

3/26/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 19:33

Date of Inspection: 02/26/2019

Serial Number: 80-000816
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/26/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 13518080A5 Exp: 08/05/2020
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.196	0.078
0.000	0.049	0.079	0.196	0.078

Number of Simulators Used: 5

Remarks:

The correct expiration date is 2-6-20 NOT 2-26-20 for the 0.05g/210L.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Signature and Printed Name GARY B FORINASH

02/26/2019
 Date