80-000822
BK

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 18:26 Date of Inspection: 12/05/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MIO 1437 2502 PITCHER / BINDER
Signature and Printed Name

12/05/2019
Date

FDLE/ATP Form 40 – March 2004

80-000822
PK

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 01:19 Date of Inspection: 11/09/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.199	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MLO 1457 0502 PITCHER / BINDER
Signature and Printed Name

11/09/2019
Date

FDLE/ATP Form 40 – March 2004

80-000822
BK

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 23:08 Date of Inspection: 09/25/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I ~~certify~~ that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D5 F. C 0502 PITCHER / BINDER
Signature and Printed Name

09/25/2019
Date

FDLE/ATP Form 40 – March 2004

80-000822
BK

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 21:59 Date of Inspection: 09/25/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 5

Remarks:
Int Det:REMIXED SOLUTION,REPLACED SIMULATOR. Non-compliance:BAD SIMULATOR JAR, REPLACED.

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MD 1437 DS 2502 PITCHER / BINDER
Signature and Printed Name

09/25/2019
Date

FDLE/ATP Form 40 – March 2004

80-000822
BK

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 15:16
Date of Inspection: 08/23/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.079	0.196	0.080

Number of Simulators Used: 5

Remarks:
Time-Date changed.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MID 1937 D/S 0502 PITCHER / BINDER
Signature and Printed Name

08/23/2019
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

7/10/19 JB

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 21:46 Date of Inspection: 07/03/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049	0.077	0.180 / 0.179	
0.000	0.049	0.078	0.178 / 0.180	
0.000	0.049	0.079	0.178 / 0.180	

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance;REPLACED SIM CONNECTORS, Cont. Non-compliance:.20 CONTROL TEST OUT OF TOLERANCE

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BINDER / PITCHER
Signature and Printed Name

07/03/2019
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

7/10/19

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 20:55 Date of Inspection: 07/03/2019 Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.044 / 0.047	0.077	0.175 / 0.181	
0.000	0.046 / 0.048	0.078	0.178 / 0.183	
0.000	0.047 / 0.048	0.078	0.182 / 0.183	

Number of Simulators Used: 5

Remarks:
05: Control Outside Tolerance/CHECKED TEMP. AND CONNECTIONS. 20: Control Outside Tolerance/CHECK TEMP. AND CONNECTIONS, . Non-compliance:0.20 OUT OF COMPLIANCE.

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

2502 1437 BINDER / PITCHER
Signature and Printed Name

07/03/2019
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

6/21/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 23:13 Date of Inspection: 06/19/2019
Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.047	0.077	0.187 / 0.194	0.080
0.000	0.047	0.078	0.192 / 0.196	0.080
0.000	0.047	0.078	0.194 / 0.197	0.080

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance/CHECKED TEMP/RECONNECTED SIM.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M10 T437 P 2502 BINDER / PITCHER
Signature and Printed Name

06/19/2019
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

5/20/19 JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O. Serial Number: 80-000822
Time of Inspection: 00:21 Date of Inspection: 05/16/2019 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.047	0.076	0.000 / 0.196	0.080
0.000	0.048	0.078	0.193 / 0.197	0.080
0.000	0.048	0.078	0.195 / 0.197	0.080

Number of Simulators Used: 5

Remarks:
20: Control Outside Tolerance/CONNECTED HOSE FROM SIMULATOR.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

2502 1437 BINDER / PITCHER
Signature and Printed Name

05/16/2019
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

4/12/19 JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 23:27 Date of Inspection: 04/10/2019 Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1099238 Exp: 03/01/2022
0.000	0.048	0.078	0.196	0.088 / 0.081
0.000	0.048	0.079	0.197	0.085 / 0.080
0.000	0.049	0.078	0.197	0.084 / 0.081

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance/ RE-SEATED DRY GAS BOTTLE .

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

1937 2502 PITCHER / BINDER
Signature and Printed Name

04/10/2019
Date

FDLE/ATP Form 40 – March 2004