

Florida Department of Law Enforcement Alcohol Testing Program

6/21/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 23:13

Date of Inspection: 06/19/2019

Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.047	0.077	0.187 / 0.194	0.080
0.000	0.047	0.078	0.192 / 0.196	0.080
0.000	0.047	0.078	0.194 / 0.197	0.080

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance/CHECKED TEMP/RECONNECTED SIM.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MIA [Signature]

R R 2502

BINDER / PITCHER
Signature and Printed Name

06/19/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/20/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 00:21

Date of Inspection: 05/16/2019

Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.047	0.076	0.000 / 0.196	0.080
0.000	0.048	0.078	0.193 / 0.197	0.080
0.000	0.048	0.078	0.195 / 0.197	0.080

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance/CONNECTED HOSE FROM SIMULATOR.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 2502 *[Signature]* 1437 BINDER / PITCHER
Signature and Printed Name

05/16/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

4/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY S.O.
Time of Inspection: 23:27

Date of Inspection: 04/10/2019

Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1099238 Exp: 03/01/2022
0.000	0.048	0.078	0.196	0.088 / 0.081
0.000	0.048	0.079	0.197	0.085 / 0.080
0.000	0.049	0.078	0.197	0.084 / 0.081

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance/ RE-SEATED DRY GAS BOTTLE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M/D [Signature]
[Signature] 2502
 PITCHER / BINDER

Signature and Printed Name

04/10/2019
Date