

Florida Department of Law Enforcement Alcohol Testing Program

5/20/19
JA

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 00:21

Date of Inspection: 05/16/2019

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.080	0.198	0.077
0.000	0.048	0.080	0.199	0.077
0.000	0.049	0.080	0.199	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MS R. 7 2582 MLO [Signature] 1437 BINDER / PITCHER
Signature and Printed Name

05/16/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

4/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 23:40

Date of Inspection: 04/10/2019

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 970670 Exp: 06/07/2021
0.000	0.048	0.080	0.197	0.078
0.000	0.048	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MID
[Signature]
1937

[Signature] 2502

PITCHER / BINDER

Signature and Printed Name

04/10/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

BK
80-000823

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO

Time of Inspection: 20:05

Date of Inspection: 02/13/2019

Serial Number: 80-000823

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.050	0.081	0.199	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MIP 1437

BINDER 2502

BINDER / PITCHER

Signature and Printed Name

02/13/2019
Date