

Florida Department of Law Enforcement Alcohol Testing Program

4/16/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Serial Number: 80-000824

Time of Inspection: 10:32

Date of Inspection: 04/13/2019

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A2 Exp: 02/05/2020
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

John G. Sommersdorf

JOHN G. SOMMERSDORF

Signature and Printed Name

04/13/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/18/19
JG

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.
Time of Inspection: 12:05

Date of Inspection: 03/16/2019

Serial Number: 80-000824
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A2 Exp: 02/05/2020
0.000	0.047	0.077	0.193	0.080
0.000	0.047	0.078	0.194	0.080
0.000	0.048	0.078	0.195	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOHN G. SOMMERSDORF

Signature and Printed Name

03/16/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

TSK
80-000824

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.
Time of Inspection: 23:06

Date of Inspection: 02/21/2019

Serial Number: 80-000824
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802H Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A2 Exp: 02/05/2020
0.000	0.047	0.076	0.194	0.080
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.196	0.080

Number of Simulators Used: 4

Remarks:

Int. Det.: LOOSE SIMULATOR JAR.
Did not detect interference on Post Test, Tightened jar 2 successful after Retest!

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

JOHN G SOMMERSDORF

Signature and Printed Name

02/21/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

1/29/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEESBURG P.D.

Time of Inspection: 00:25

Date of Inspection: 01/20/2019

Serial Number: 80-000824

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34917080A2 Exp: 02/05/2020
0.000	0.044 / 0.047	0.076	0.193	0.079
0.000	0.044 / 0.048	0.076	0.194	0.079
0.000	0.045 / 0.048	0.077	0.195	0.080

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance. REPLACE O-RING, RETIGHTEN JAR

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Signature and Printed Name JOHN E. SOMMERSDORF

Date 01/20/2019