

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHATTAHOOCHEE PD
Time of Inspection: 13:31

Date of Inspection: 11/29/2019

Serial Number: 80-000887
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.079 / 0.050	0.080	0.195	0.079
0.000	0.079 / 0.050	0.080	0.196	0.080
0.000	0.080 / 0.050	0.080	0.197	0.080

Number of Simulators Used: 5


Remarks:

A F / M A: . 05: Control Outside Tolerance.

Used .08 simulator

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



FREDERICK C OBRIAN

Signature and Printed Name

11/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHATTAHOOCHEE PD
Time of Inspection: 13:51

Date of Inspection: 05/29/2019

Serial Number: 80-000887
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021
0.000	0.049	0.078	0.188 / 0.197	0.080
0.000	0.049	0.078	0.189 / 0.197	0.080
0.000	0.049	0.078	0.189 / 0.197	0.079


Number of Simulators Used: 5

Remarks:

Int Det:TIGHTEN SIM CAP. 20: Control Outside Tolerance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name FREDERICK C O'BRIAN

05/29/2019
 Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 30 2019

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHATTAHOOCHEE PD
Time of Inspection: 17:38

Date of Inspection: 04/25/2019

Serial Number: 80-000887
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 970670 Exp: 06/07/2021
0.000	0.047	0.076	0.193	0.080
0.000	0.048	0.076	0.194	0.080
0.000	0.049	0.076	0.194	0.080

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

FCO Obrian

FREDERICK C OBRIAN

Signature and Printed Name

04/25/2019
Date