

946
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 07:27

Date of Inspection: 04/29/2019

Serial Number: 80-000946
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 10818080A5 Exp: 06/05/2020
/ 0.000	0.049	0.079	0.197	0.078
/ 0.000	0.049	0.079	0.196	0.078
/ 0.000	0.049	0.079	0.197	0.078

Number of Simulators Used: 5

Remarks:
00: Ambient Fail

Ambient Alcohol in Air During test
[Signature]

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] 3571
Signature and Printed Name MICHAEL P KERN

04/29/2019
Date