

996 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE P.D.

Time of Inspection: 01:47

Date of Inspection: 12/30/2019

Serial Number: 80-000996

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A2 Exp: 12/05/2021
0.000	0.045	0.076	0.181 / 0.198	0.081
0.000	0.048	0.077	0.195 / 0.199	0.081
0.000	0.048	0.077	0.198 / 0.199	0.081

Number of Simulators Used: 4


Remarks:

20: Control Outside Tolerance.

* Nose bent

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 JASON M SANCHEZ
Signature and Printed Name

12/30/2019
Date

9910 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE P.D.
Time of Inspection: 21:17

Date of Inspection: 02/27/2019

Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:29017080A2 Exp: 12/05/2019
0.000	0.046	0.075	0.194	0.080
0.000	0.046	0.076	0.196	0.080
0.000	0.046	0.076	0.197	0.079

Number of Simulators Used: 4

Remarks: After further review it was determined the 20 solution was expired.

The above instrument complies (^{SS}) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jason M Sanchez 722 JASON M SANCHEZ
Signature and Printed Name

02/27/2019
Date