

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.
Time of Inspection: 23:56

Date of Inspection: 05/29/2019

Serial Number: 80-000999
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:20802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201802B Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:915310 Exp: 03/07/2021
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.079

Number of Simulators Used: 5

Remarks:
NONE

05 Exp date 02/06/2020 correct lot# 201802B KB 5/29/2019
08 correct lot# 201709A KB 05/29/2019

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KEITH D BETHAM

Signature and Printed Name

05/29/2019
Date