

1004
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 14:44

Date of Inspection: 05/27/2019

Serial Number: 80-001004
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1981708GA3 Exp: 09/05/2019
0.000	0.045 / 0.047	0.078	0.197	0.079
0.000	0.044 / 0.048	0.079	0.198	0.079
0.000	0.044 / 0.047	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:


05: Control Outside Tolerance.

During the first 0.05g/210L Test, "Control Outside Tolerance" notification was observed. Troubleshooting was conducted. I ensured the simulator container lid was tightened. A follow up test was conducted with no issues observed.


Dep. Robin, G. #3877

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Dep. Robin, G. #3877
Signature and Printed Name

05/27/2019
Date

1004 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 19:25

Date of Inspection: 03/28/2019

Serial Number: 80-001004
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19817080A3 Exp: 09/05/2019
0.000	0.050	0.080	0.000 / 0.198	0.077
0.000	0.049	0.080	0.000 / 0.198	0.078
0.000	0.050	0.080	0.000 / 0.198	0.078

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

Rim of Simulator has a crack in the glass. Replaced glass jar and completed inspection w/no issues.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas A Garrity

THOMAS A GARRITY

Signature and Printed Name

03/28/2019
Date