

Florida Department of Law Enforcement Alcohol Testing Program

2/8/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 12:13

Date of Inspection: 02/04/2019

Serial Number: 80-001005
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG815602 Exp: 06/05/2020
0.000	0.045	0.071 / 0.076	0.192	0.079
0.000	0.047	0.074 / 0.077	0.194	0.078
0.000	0.048	0.076 / 0.077	0.195	0.078


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance UNK.PROB. REPEAT OK..

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


M DEANE
Signature and Printed Name

02/04/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/8/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 21:16

Date of Inspection: 02/02/2019

Serial Number: 80-001005
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG815602 Exp: 06/05/2020
0.000	0.049	0.078	0.195	0.078
0.000	0.050	0.078	0.197	0.078
0.000	0.050	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: ^{MD FAIL} TOO MUCH MAS INTRODUCED REPEAT ~~OK~~.
SPOKE TO FDLE/ATP - REDO NEXT DAY - OK

2/8/19
JD

The above instrument complies (^{MD} ~~X~~) does not comply (^{MD} X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Deane
M DEANE
Signature and Printed Name

02/02/2019
Date