

Florida Department of Law Enforcement Alcohol Testing Program

5/21/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 09:30

Date of Inspection: 05/21/2019

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:906364 Exp: 02/07/2021
0.000	0.051	0.081	0.202	0.082
0.000	0.051	0.080	0.202	0.082
0.000	0.052	0.081	0.202	0.083

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpt. WEBER #896

RAY M WEBER
Signature and Printed Name

05/21/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/2/19
J/O

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 09:39

Date of Inspection: 04/29/2019

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802E Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:906364 Exp: 02/07/2021
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.081	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER #896

RAY M WEBER
Signature and Printed Name

04/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/29/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 08:31

Date of Inspection: 03/28/2019

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.052	0.083	0.201	0.080
0.000	0.052	0.083	0.201	0.080
0.000	0.052	0.083	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

gpl. WEBER #896
RAY M WEBER

 Signature and Printed Name

03/28/2019
 Date

Florida Department of Law Enforcement Alcohol Testing Program

2/20/19
JA

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 06:40

Date of Inspection: 02/20/2019

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.050	0.079	0.200	0.082
0.000	0.051	0.080	0.200	0.081
0.000	0.052	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER #896

RAY M WEBER
Signature and Printed Name

02/20/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/12/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 08:24

Date of Inspection: 01/28/2019

Serial Number: 80-001007
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:684131 Exp: 01/23/2020
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL RAY WEBER #596

RAY M WEBER

Signature and Printed Name

01/28/2019
Date