

# Florida Department of Law Enforcement Alcohol Testing Program

4/11/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 15:18

Date of Inspection: 03/28/2019

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:686331 Exp: 01/31/2020
/ 0.000	0.047	0.076	0.193	0.078
/ 0.000	0.048	0.076	0.194	0.078
/ 0.000	0.048	0.078	0.195	0.078

Number of Simulators Used: 4

**Remarks:**

00: Ambient Fail UNKN WHAT CAUSED.

Possible RESIDUAL Mouth Alcohol liquid that was spilled on floor near TRASH CAN PRIOR to AIR PARGE.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

\_\_\_\_\_  
 Signature and Printed Name

03/28/2019  
 Date