

Florida Department of Law Enforcement Alcohol Testing Program

BK
80-001053

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NASSAU CO SO
Time of Inspection: 16:05

Date of Inspection: 04/01/2019

Serial Number: 80-001053
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A5 Exp: 10/05/2019
0.000	0.049	0.079	0.195	0.081
0.000	0.049	0.080	0.195	0.081
0.000	0.049	0.080	0.196	0.081

Number of Simulators Used: 4

Remarks:
TESTED IMP SAMPLE.

Amended lot 201802B
5-28-19

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Terry L Nye

TERRY L NYE

Signature and Printed Name

04/01/2019
Date

80-001053
BK

Kirkland, Brett

From: Kirkland, Brett
Sent: Wednesday, May 22, 2019 9:15 AM
To: tnye@nassauso.com
Subject: ATP: 80-001053 4/1/19 Agency Inspection

Terry,

I was reviewing the agency inspection for instrument 80-001053 for April and noticed that there is an incomplete lot # listed. For the 0.05 solution, it is written as "201802". Is this supposed to be 201802B? Please send in an hand-written amended Form 40 so that we can have it on file.

Thanks

Brett Kirkland, Ph.D.

Program Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office: (850) 617-1275
Cell: (850) 631-1351

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

80-001053
BK

Kirkland, Brett

From: Kirkland, Brett
Sent: Tuesday, May 28, 2019 11:56 AM
To: 'Terry Nye'
Subject: RE: amended

Received. Thank you

Brett

From: Terry Nye [mailto:tnye@nassauso.com]
Sent: Tuesday, May 28, 2019 11:51 AM
To: Kirkland, Brett
Subject: amended

*Nassau County Sheriff's Office
Corrections Lieutenant Terry Nye
Office 904-548-4077
Cell 904-583-9108*