

Amended

Florida Department of Law Enforcement Alcohol Testing Program

*5/2/19
JA*

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 09:56

Date of Inspection: 04/28/2019

Serial Number: 80-001055
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22817080A4 Exp: 10/05/2019
0.000	0.049	0.080	0.200	0.000 / 0.080
0.000	0.050	0.079	0.200	0.000 / 0.080
0.000	0.050	0.080	0.201	0.000 / 0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Dry gas cylinder wasn't tighter

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TEAQUILIA L HAMMONDS

TEAQUILIA L HAMMONDS

Signature and Printed Name

04/28/2019
Date