

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

3/27/19  
JP

Agency: NEW PORT RICHEY PD  
Time of Inspection: 21:33

Date of Inspection: 03/26/2019

Serial Number: 80-001064  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.000 / 0.000		
0.000	0.047	0.000 / 0.000		
0.000	0.047	0.000 / 0.000		

Number of Simulators Used: 4

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

SIMULATOR JAR FOR .08 SOLUTION HAD CRACKED AND  
PIECE BROKE OFF AT THREADS ALONG THE SEAL,  
REPLACED WITH NEW JAR AND DID INSPECTION AGAIN.

The above instrument complies (       ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt. J. Engel NP407  JASON S ENGEL  
Signature and Printed Name

03/26/2019  
Date