

80-001069  
BK

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 13:34

Date of Inspection: 01/22/2019

Serial Number: 80-001069  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 05218080A3 Exp: 05/05/2020
0.000 / 0.000	0.050	0.080	0.203	0.083
/ 0.000	0.051	0.080	0.198	0.083
/ 0.000	0.050	0.080	0.201	0.083

Number of Simulators Used: 5

Remarks:  
00: RFI Detect . Int Det: RFI Detect.

### Amended Inspection:

I received an RFI during Alcohol Free portion of this inspection.  
I searched the immediate area but was unable to locate the source of the RFI.  
I re-ran the Alcohol Free portion of the inspection without further issue.  
I received another RFI during the Int. Detect portion of this inspection.  
I expanded the search area and located a wireless laptop being used directly adjacent to the breath testing room in the sally-port.  
After clearing this possible RFI source, I finished the remainder of this inspection without further issue.

Sgt. M Potter 7599 1.22.2019

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL R POTTER

Signature and Printed Name

01/22/2019  
Date