

Florida Department of Law Enforcement Alcohol Testing Program

80-001074

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 15:20

Date of Inspection: 11/20/2019

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000	0.049	0.078	0.157 / 0.193	0.078
0.000	0.049	0.079	0.173 / 0.194	0.078
0.000	0.049	0.079	0.179 / 0.194	0.078

Number of Simulators Used: 4

Remarks:

Time-Date changed. 20: Control Outside Tolerance CHANGED SIM

Change simulator the report

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

3922
BARRY NIXON

Signature and Printed Name

11/20/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001074
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 14:38

Date of Inspection: 10/22/2019

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22419080A3 Exp: 10/05/2021
0.000	0.049	0.078	0.186 / 0.196	0.079
0.000	0.049	0.078	0.195 / 0.197	0.079
0.000	0.050	0.078	0.196 / 0.198	0.079


Number of Simulators Used: 1

Remarks:

20: Control Outside Tolerance TIGHTEN SIM THEN REPEAT.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BARRY NIXON
Signature and Printed Name

10/22/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

80-001074
BK

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 16:27

Date of Inspection: 08/20/2019

Serial Number: 80-001074
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A3 Exp: 03/05/2020
0.000	0.043 / 0.047	0.077	0.195	0.078
0.000	0.046 / 0.047	0.078	0.197	0.079
0.000	0.046 / 0.047	0.078	0.198	0.079


Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. *changed simulator for the 0.05g/210L*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BARRY NIXON
Signature and Printed Name

08/20/2019
Date