

1077
88

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 17:48

Date of Inspection: 12/31/2019

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1076482 Exp: 01/15/2022
0.000 / 0.000	0.048	0.078	0.197	0.077
0.000 / 0.000	0.048	0.079	0.198	0.077
0.000 / 0.000	0.049	0.080	0.198	0.077

Number of Simulators Used: 5

Remarks:
00: Ambient Fail. Reconnected the Alcohol Free simulator. Lost Power during post diagnostic checks. Restored Power and conducted another Diagnostic checks. See second page

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GARY W. CARTWRIGHT JR.
Signature and Printed Name

12/31/2019
Date

1077 SP

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001077
12/31/2019
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

1077
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 15:08

Date of Inspection: 01/11/2019

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:980558 Exp: 07/09/2021
0.000	0.046	0.077	0.000 / 0.194	0.078
0.000	0.047	0.078	0.000 / 0.196	0.078
0.000	0.048	0.078	0.000 / 0.196	0.079

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. Did not have simulator connected to intoxilyzer on first 0.20 test. *[Signature]*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

GARY W. CARTWRIGHT JR
Signature and Printed Name

01/11/2019
Date