

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 10:57

Date of Inspection: 12/20/2019

Serial Number: 80-001083  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.048	0.077	0.212 / 0.195	0.080
0.000	0.049	0.079	0.208 / 0.198	0.080
0.000	0.049	0.079	0.206 / 0.198	0.079

Number of Simulators Used: 5

Remarks:  
20: Control Outside Tolerance. INSPECTED SIM AND RERAN TEST

*AMENDED INSPECTION*  
INSPECTION FOUND .200 SIMULATOR TEMPERATURE HAD EXCEEDED 34C (H-20) CAUSING HIGH READINGS. I TURNED OFF SIMULATOR AND ALLOWED THE SIMULATOR TO COOL. AFTER COOLING I TURNED SIMULATOR BACK ON AND ALLOWED THE SIMULATOR TO RETURN TO OPERATING TEMPERATURE. AFTER SIMULATOR RETURNED TO PROPER OPERATING TEMPERATURE I RERAN TEST WITH NO FURTHER ISSUES.

AGENCY INSPECTOR  
*R. D. LIGHTSEY* 6906 JACKSONVILLE STATION OFFICE 12/20/2019 11:56

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*

ROBERT D LIGHTSEY  
Signature and Printed Name

12/20/2019  
Date