

# Florida Department of Law Enforcement Alcohol Testing Program

6/28/19  
JD

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 15:45

Date of Inspection: 06/13/2019

Serial Number: 80-001139  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 970670 Exp: 06/07/2021
0.000	0.047	0.196 / 0.076	0.197	0.080
0.000	0.047	0.198 / 0.078	0.199	0.080
0.000	0.048	0.199 / 0.078	0.199	0.080


Number of Simulators Used: 5

**Remarks:**

(200)  
08: Control Outside Tolerance. WRONG SIMULATOR ATTACHED DURING .08. TEST CONDUCTED A SECOND TIME WITH .08 REFERENCE SOLUTION.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 YVETTE D WASHINGTON  
 Signature and Printed Name

06/13/2019  
 Date