

1150  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 08:17

Date of Inspection: 03/25/2019

Serial Number: 80-001150  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) <del>201811C</del> 201808E Lot#: <del>201811C</del> Exp: <del>11/13/2020</del> 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 840875 Exp: 09/19/2020
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079


Number of Simulators Used: 5

Remarks:

Lot# for 0.08g/210L test was incorrect, Correct lot# was 201808E, EXPIRATION OF 08/22/2020

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

03/25/2019  
Date