

Florida Department of Law Enforcement Alcohol Testing Program

5/21/19
JW

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 09:21

Date of Inspection: 05/21/2019

Serial Number: 80-001151
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.203 | 0.083 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.083 |
| 0.000 | 0.050 | 0.081 | 0.204 | 0.083 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER #896
RAY M WEBER

 Signature and Printed Name

05/21/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

5/2/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 09:33

Date of Inspection: 04/29/2019

Serial Number: 80-001151
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.079 | 0.201 | 0.082 |
| 0.000 | 0.049 | 0.079 | 0.201 | 0.082 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.083 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RAY M WEBER #596

RAY M WEBER

Signature and Printed Name

04/29/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

3/29/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 08:26

Date of Inspection: 03/28/2019

Serial Number: 80-001151
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.201 | 0.083 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.083 |
| 0.000 | 0.050 | 0.081 | 0.202 | 0.083 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


CPL. WEBER #896
RAY M WEBER

 Signature and Printed Name

03/28/2019
 Date

Florida Department of Law Enforcement Alcohol Testing Program

2/20/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 06:34

Date of Inspection: 02/20/2019

Serial Number: 80-001151
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#: 201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 970670 Exp: 06/07/2021 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.079 | 0.201 | 0.083 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.083 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.082 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


CPL. RAY WEBER #596
RAY M WEBER

 Signature and Printed Name

02/20/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD
Time of Inspection: 08:18

Date of Inspection: 01/28/2019

Serial Number: 80-001151
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020 | 0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020 | 0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:970670 Exp: 06/07/2021 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.202 | 0.084 |
| 0.000 | 0.050 | 0.080 | 0.202 | 0.084 |
| 0.000 | 0.051 | 0.080 | 0.202 | 0.084 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gal. KAYWEBER #896

RAY M WEBER

Signature and Printed Name

01/28/2019
Date