

1166 SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY SO  
Time of Inspection: 14:54

Date of Inspection: 05/27/2019

Serial Number: 80-001166  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19817080A3 Exp: 09/05/2019
0.000	0.047 / 0.045	0.076	0.197 / 0.198	0.082
0.000	0.047 / 0.050	0.077	0.199 / 0.198	0.082
0.000	0.048 / 0.050	0.078	/ 0.198	0.083

Number of Simulators Used: 5

Remarks:

05: Ambient Fail. 20: Ambient Fail.

*Ambient fail notifications were observed during inspection. Troubleshooting and area was cleared for any potential influences. A follow up test was conducted with no issues observed.*

*[Signature]*  
Det. Robin, G. #3877

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]* Det. Robin, G. #3877 GASTON A ROBIN  
Signature and Printed Name

05/27/2019  
Date