

Florida Department of Law Enforcement Alcohol Testing Program

5/2/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 00:26

Date of Inspection: 04/30/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
/ 0.000	0.049	0.078	0.196	0.080
/ 0.000	0.048	0.078	0.200	0.079
/ 0.000	0.048	0.078	0.196	0.080

Number of Simulators Used: 4

Remarks:
00: Ambient Fail.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DEBORAH REASONER

 Signature and Printed Name

04/30/2019
 Date

Graham, Thomas

From: Graham, Thomas
Sent: Thursday, May 02, 2019 7:32 AM
To: Alcohol Testing Program
Subject: RE: Clermont April inspection

80-001174
5/2/19
JR

Sgt. Reasoner,

Have you been able to upload your instrument to FDLE? Also, please remember to make a note as to why you believe you received an ambient fail message during your inspection on the Form 40.

Thank you.

TJ

From: Alcohol Testing Program
Sent: Tuesday, April 30, 2019 8:15 AM
To: Graham, Thomas
Subject: FW: Clermont April inspection

From: Deborah Reasoner [<mailto:DReasoner@clermontfl.org>]
Sent: Tuesday, April 30, 2019 4:08 AM
To: Chambers, Rebecca; Bell, Danielle; Alcohol Testing Program; winterspringsbar@flhsmv.gov; wsievers@sao5.org
Subject: Clermont April inspection

Debbie Reasoner
Police Sergeant
Clermont Police Department
Bravo Squad
DRE/THI

3600 US 27., Clermont, FL 34711
Tel 352-536-8461 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



Florida Department of Law Enforcement Alcohol Testing Program

3/29/19
JO

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 03:00

Date of Inspection: 03/29/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.049	0.078	0.194	0.079
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.079	0.197	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


141
DEBORAH REASONER

 Signature and Printed Name

03/29/2019
 Date

Florida Department of Law Enforcement Alcohol Testing Program

2/27/19
JE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 02:42

Date of Inspection: 02/25/2019

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.076	0.195	0.079
0.000	0.048	0.076	0.196	0.079
0.000	0.048	0.076	0.195	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt Deborah Reasoner DEBORAH REASONER
Signature and Printed Name

02/25/2019
Date

Florida Department of Law Enforcement Alcohol Testing Program

2/12/19
JD

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 23:25

Date of Inspection: 01/27/2019

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02618080A2 Exp: 03/05/2020
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

141

DEBORAH REASONER

Signature and Printed Name

01/27/2019
Date