

1253
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 19:59

Date of Inspection: 08/19/2019

Serial Number: 80-001253
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

Non-compliance: .

At the time of the inspection, there was a "Diagnostic Fail" notification. I inspected the outside of instrument with no apparent discrepancies. A follow up Diagon^{al} Diagnostics Test was conducted with no faults observed. See instrument print out for further details. A follow up Inspection Test was conducted and determined within compliance.

[Signature]
Gp. Robin, G. #3877

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] Gp. Robin, G. #3877
Signature and Printed Name GASTON A ROBIN

08/19/2019
Date

1253
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 14:57

Date of Inspection: 05/27/2019

Serial Number: 80-001253
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19817080A3 Exp: 09/05/2019
0.000	0.049	0.081	0.199	0.080
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.081	0.200	0.079

Number of Simulators Used: 5

Remarks:


A F / M A: .

During the Alcohol Free / Mouth Alcohol Test, a malfunction was made. Troubleshooting was conducted. A follow up test was conducted with no issues observed.


Dep. Robin, G. #3877

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Dep. Robin, G. #3877
Signature and Printed Name
GASTON A ROBIN

05/27/2019
Date

1253
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 18:40

Date of Inspection: 03/28/2019

Serial Number: 80-001253
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.050	0.080	0.000 / 0.000	
0.000	0.051	0.080	0.000 / 0.000	
0.000	0.050	0.080	0.000 / 0.000	

Number of Simulators Used: 5

Remarks:

Time-Date changed. 20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:

RIM OF SIMULATOR HAS A CRACK IN THE GLASS. REPLACED GLASS JAR AND CONDUCTED ANOTHER INSPECTION. NO ISSUES ON REINSPECTION.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Thomas A. Garrity

THOMAS A GARRITY

Signature and Printed Name

03/28/2019
Date