

1264
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.
Time of Inspection: 14:35

Date of Inspection: 05/27/2019

Serial Number: 80-001264
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201802G Exp: 02/22/2020	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19817080A3 Exp: 09/05/2019
0.000	0.029 / 0.048	0.076	0.187 / 0.196	0.080
0.000	0.029 / 0.048	0.078	0.192 / 0.196	0.081
0.000	0.028 / 0.048	0.079	0.194 / 0.197	0.081

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

During the first 0.05g/210L Test and 0.20g/210L Test, "Control Outside Tolerance" notifications were observed. Troubleshooting was conducted. I ensured the simulator container lids were tightened. A follow up test was conducted with no issues observed.

[Signature]
Det. Robin, G. #3877

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] Det. Robin, G. #3877 GASTON A ROBIN
Signature and Printed Name

05/27/2019
Date